

CORNING GOSSELIN SAS  
123 Rue de Caëstre - CS40019  
59529 HAZEBROUCK CEDEX  
BORRE - FRANCE

Date of certificate : 2021 - 05 - 25

Page : 1/1

### **Product description**

*Container 60mL PP RedScrewCap Sterile 70/Bag 700/CS*

### **Catalogue number**

*TP35C-002*

### **Manufacturing date**

*2021/04/21*

### **Lot number**

*785135*

### **Sterility Assurance Level**

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10<sup>-3</sup>

### **Expiration date**

*2024/04/01*

### **Expiration statement**

*This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.*

### **Performance testing (AQL)**

*Inspection based on a statistically significant number of samples towards internal or agreed specifications that determine Assurance Quality Level.*



Cédric MAES  
Plant Quality Manager, Borre  
maesc@corning.com