

CORNING GOSSELIN SAS
123 Rue de Caëstre - CS40019
59529 HAZEBROUCK CEDEX
BORRE - FRANCE

Date of certificate : 2021 - 08 - 12

Page : 1/1

Product description

Container 60mL PP RedScrewCap Sterile 70/Bag 700/CS

Catalogue number

TP35C-002

Manufacturing date

2021/06/10

Lot number

787656

Sterility Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10⁻³

Expiration date

2024/06/01

Expiration statement

This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards internal or agreed specifications that determine Assurance Quality Level.



Hugues Vanderschelden
Plant Quality Manager, Borre
vandersch2@corning.com