

CORNING GOSSELIN SAS  
123 Rue de Caëstre - CS40019  
59529 HAZEBROUCK CEDEX  
BORRE - FRANCE  
[www.corning.com/lifesciences](http://www.corning.com/lifesciences)  
Refer to website for regional contact information

Date of certificate : 2022 - 10 - 18

Page 1

### **Product description**

*Container 60mL PP WhiteScrewCap 70/ Bag 700/CS*

### **Catalogue number**

*TP35C-013*

### **Manufacturing date**

*2022/08/30*

### **Lot number**

*829782*

**Quality Management System** : Complies with the current version of the ISO9001 standard

### **Sterility Assurance Level**

Non sterile

### **Expiration statement**

*N/A*

### **Performance testing (AQL)**

*Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.*

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release



Hugues Vanderschelden  
Plant Quality Manager, Borre  
[vandersch2@corning.com](mailto:vandersch2@corning.com)